

CE DECLARATION OF CONFORMITY

Equipments Types :

TR2510B (M3.UB) 10A/1.5V. (Fluke P/N 4379640)
TR2500B (S.UE) 500A/1.5V (Fluke P/N 4379691)

Manufacturer: **UNIVERSAL TECHNIC**
Address: **1, rue Robert et Sonia DELAUNAY**
75011 PARIS FRANCE

European Standards:

EMC EN50082-1: 1992 Generic Immunity Standard.
Part 1. Residential, commercial, and light industry.

EN50081-1: 1992 Generic Emission Standard.
Part 1. Residential, commercial, and light industry.

Safety IEC 61010-1, Ed.2 (2010),
Safety requirements for electrical equipment for measurement,
Control and laboratory use:
Part 1: General Requirements.
IEC 61010-2-031, Ed.1 (2008).
Part 2: Particular requirements for hand-held current probe assemblies
for electrical measurement and test.
IEC 61010-2-032 Ed.1 (2012).
Part 2: Particular requirements for hand-held current clamps for
electrical test and measurement.

Description of Equipment: AC Clamp On Current Probes.

I certify that the apparatus identified above conforms to the requirements of Council Directives:

- (1) Electromagnetic Compatibility Directive 89/336/EEC
- (2) Low Voltage Directive 73/23/EEC
- (3) CE Marking Directive 93/68/EEC

Signed:

Name: **Valéry Le Page**
Managing Director



Date: 24/04/2025

Company Name: **UNIVERSAL TECHNIC**
1 Rue Robert et Sonia DELAUNAY
75011 PARIS – France.