

# EU Declaration of Conformity

Document number CE02076

We GMC-I PROSYS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH  
Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 3000XLB24	P-02.716.8	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

## Low Voltage Directive (LVD) 2014/35/EU

### EN 61010-1:2010+A1:2019

Safety requirements for electrical equipment for measurement, control, and laboratory use.

600V CAT IV  
1000V CAT III  
Pollution Degree 2

### EN IEC 61010-2-032:2021+A11:2021

Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

Type B

## EMC Directive 2014/30/EU

### EN 61326-2-2:2013


EMC requirements. Particular requirements. Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems.

Susceptibility Criteria A  
Emission Class B

## RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSYS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature		Name	Robin Brooks
Position	R&D Director	Date	1st April 2025

# EU Declaration of Conformity

Document number CE02077

We GMC-I PROSYS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH  
 Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 3000XLB36	P-02.716.9	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

## Low Voltage Directive (LVD) 2014/35/EU

### EN 61010-1:2010+A1:2019

Safety requirements for electrical equipment for measurement, control, and laboratory use.

600V CAT IV  
 1000V CAT III  
 Pollution Degree 2

### EN IEC 61010-2-032:2021+A11:2021

Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

Type B

## EMC Directive 2014/30/EU

### EN 61326-2-2:2013


EMC requirements. Particular requirements. Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems.

Susceptibility Criteria A  
 Emission Class B

## RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSYS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature		Name	Robin Brooks
Position	R&D Director	Date	1st April 2025

# EU Declaration of Conformity

Document number CE02078

We GMC-I PROSYS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH  
 Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 3000XLB48	P-02.717.0	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

## Low Voltage Directive (LVD) 2014/35/EU

### EN 61010-1:2010+A1:2019

Safety requirements for electrical equipment for measurement, control, and laboratory use.

600V CAT IV  
 1000V CAT III  
 Pollution Degree 2

### EN IEC 61010-2-032:2021+A11:2021

Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

Type B

## EMC Directive 2014/30/EU

### EN 61326-2-2:2013


EMC requirements. Particular requirements. Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems.

Susceptibility Criteria A  
 Emission Class B

## RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSYS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature		Name	Robin Brooks
Position	R&D Director	Date	1st April 2025

# EU Declaration of Conformity

Document number CE02079

We GMC-I PROSYS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH  
Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 3003XLB24	P-02.717.3	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

## Low Voltage Directive (LVD) 2014/35/EU

### EN 61010-1:2010+A1:2019

Safety requirements for electrical equipment for measurement, control, and laboratory use.

600V CAT IV  
1000V CAT III  
Pollution Degree 2

### EN IEC 61010-2-032:2021+A11:2021

Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

Type B

## EMC Directive 2014/30/EU

### EN 61326-2-2:2013


EMC requirements. Particular requirements. Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems.

Susceptibility Criteria A  
Emission Class B

## RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSYS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature		Name	Robin Brooks
Position	R&D Director	Date	1st April 2025

# EU Declaration of Conformity

Document number CE02080

We GMC-I PROSYS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH  
 Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 3003XLB36	P-02.717.4	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

## Low Voltage Directive (LVD) 2014/35/EU

### EN 61010-1:2010+A1:2019

Safety requirements for electrical equipment for measurement, control, and laboratory use.

600V CAT IV  
 1000V CAT III  
 Pollution Degree 2

### EN IEC 61010-2-032:2021+A11:2021

Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

Type B

## EMC Directive 2014/30/EU

### EN 61326-2-2:2013


EMC requirements. Particular requirements. Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems.

Susceptibility Criteria A  
 Emission Class B

## RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSYS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature		Name	Robin Brooks
Position	R&D Director	Date	1st April 2025

# EU Declaration of Conformity

Document number CE02081

We GMC-I PROSYS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH  
 Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 3003XLB48	P-02.717.5	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

## Low Voltage Directive (LVD) 2014/35/EU

### EN 61010-1:2010+A1:2019

Safety requirements for electrical equipment for measurement, control, and laboratory use.

600V CAT IV  
 1000V CAT III  
 Pollution Degree 2

### EN IEC 61010-2-032:2021+A11:2021

Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

Type B

## EMC Directive 2014/30/EU

### EN 61326-2-2:2013


EMC requirements. Particular requirements. Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems.

Susceptibility Criteria A  
 Emission Class B

## RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSYS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature		Name	Robin Brooks
Position	R&D Director	Date	1st April 2025

# EU Declaration of Conformity

Document number CE02153

We GMC-I PROSYS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH  
 Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 6000XLB24	P-02.763.4	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

## Low Voltage Directive (LVD) 2014/35/EU

### EN 61010-1:2010+A1:2019

Safety requirements for electrical equipment for measurement, control, and laboratory use.

600V CAT IV  
 1000V CAT III  
 Pollution Degree 2

### EN IEC 61010-2-032:2021+A11:2021

Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

Type B

## EMC Directive 2014/30/EU

### EN 61326-2-2:2013


EMC requirements. Particular requirements. Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems.

Susceptibility Criteria A  
 Emission Class B

## RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSYS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature		Name	Robin Brooks
Position	R&D Director	Date	1st April 2025

# EU Declaration of Conformity

Document number CE02154

We GMC-I PROSYS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH  
Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 6000XLB36	P-02.763.5	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

## Low Voltage Directive (LVD) 2014/35/EU

### EN 61010-1:2010+A1:2019

Safety requirements for electrical equipment for measurement, control, and laboratory use.

600V CAT IV  
1000V CAT III  
Pollution Degree 2

### EN IEC 61010-2-032:2021+A11:2021

Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

Type B

## EMC Directive 2014/30/EU

### EN 61326-2-2:2013


EMC requirements. Particular requirements. Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems.

Susceptibility Criteria A  
Emission Class B

## RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSYS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature		Name	Robin Brooks
Position	R&D Director	Date	1st April 2025

# EU Declaration of Conformity

Document number CE02155

We GMC-I PROSYS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH  
 Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 6000XLB48	P-02.763.6	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

## Low Voltage Directive (LVD) 2014/35/EU

### EN 61010-1:2010+A1:2019

Safety requirements for electrical equipment for measurement, control, and laboratory use.

600V CAT IV  
 1000V CAT III  
 Pollution Degree 2

### EN IEC 61010-2-032:2021+A11:2021

Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

Type B

## EMC Directive 2014/30/EU

### EN 61326-2-2:2013


EMC requirements. Particular requirements. Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems.

Susceptibility Criteria A  
 Emission Class B

## RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSYS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature		Name	Robin Brooks
Position	R&D Director	Date	1st April 2025

# EU Declaration of Conformity

Document number CE02156

We GMC-I PROSYS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH  
Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 6003XLB24	P-02.763.7	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

## Low Voltage Directive (LVD) 2014/35/EU

### EN 61010-1:2010+A1:2019

Safety requirements for electrical equipment for measurement, control, and laboratory use.

600V CAT IV  
1000V CAT III  
Pollution Degree 2

### EN IEC 61010-2-032:2021+A11:2021

Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

Type B

## EMC Directive 2014/30/EU

### EN 61326-2-2:2013


EMC requirements. Particular requirements. Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems.

Susceptibility Criteria A  
Emission Class B

## RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSYS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature		Name	Robin Brooks
Position	R&D Director	Date	1st April 2025

# EU Declaration of Conformity

Document number CE02157

We GMC-I PROSYS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH  
Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 6003XLB36	P-02.763.8	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

## Low Voltage Directive (LVD) 2014/35/EU

### EN 61010-1:2010+A1:2019

Safety requirements for electrical equipment for measurement, control, and laboratory use.

600V CAT IV  
1000V CAT III  
Pollution Degree 2

### EN IEC 61010-2-032:2021+A11:2021

Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

Type B

## EMC Directive 2014/30/EU

### EN 61326-2-2:2013


EMC requirements. Particular requirements. Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems.

Susceptibility Criteria A  
Emission Class B

## RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSYS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature		Name	Robin Brooks
Position	R&D Director	Date	1st April 2025

# EU Declaration of Conformity

Document number CE02158

We GMC-I PROSYS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH  
 Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 6003XLB48	P-02.763.9	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

## Low Voltage Directive (LVD) 2014/35/EU

### EN 61010-1:2010+A1:2019

Safety requirements for electrical equipment for measurement, control, and laboratory use.

600V CAT IV  
 1000V CAT III  
 Pollution Degree 2

### EN IEC 61010-2-032:2021+A11:2021

Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

Type B

## EMC Directive 2014/30/EU

### EN 61326-2-2:2013


EMC requirements. Particular requirements. Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems.

Susceptibility Criteria A  
 Emission Class B

## RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSYS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature		Name	Robin Brooks
Position	R&D Director	Date	1st April 2025