

## **EU Declaration of Conformity**

### **Document number CE06008**

We GMC-I PROSyS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 300MHB6	P-06.461.5	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

## Low Voltage Directive (LVD) 2014/35/EU EN 61010-1:2010+A1:2019 Safety requirements for electrical equipment for measurement, control, and laboratory use. 600V CAT IV 1000V CAT III Pollution Degree 2 EN IEC 61010-2-032:2021+A11:2021 Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

### EMC Directive 2014/30/EU

### EN 61326-1:2013

Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements

Susceptibility Criteria A Emission Class B

### RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSyS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature	NBOOL	Name	Robin Brooks
Position	R&D Director	Date	1st April 2025



## **EU Declaration of Conformity**

### **Document number CE06009**

We GMC-I PROSyS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 300MHB8	P-06.461.6	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

## Low Voltage Directive (LVD) 2014/35/EU EN 61010-1:2010+A1:2019 Safety requirements for electrical equipment for measurement, control, and laboratory use. 600V CAT IV 1000V CAT III Pollution Degree 2 EN IEC 61010-2-032:2021+A11:2021 Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

### EMC Directive 2014/30/EU

### EN 61326-1:2013

Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements

Susceptibility Criteria A Emission Class B

### RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSyS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature

Name

Robin Brooks



## **EU Declaration of Conformity**

### **Document number CE06010**

We GMC-I PROSyS 1 Potter Place, Skelmersdale, Lancashire WN8 9PH Certify that the product/s described below is/are in conformity with the following directive(s):

Model	Order Number	Description
DRANFLEX 300MHB12	P-06.461.7	Flexible Rogowski Probe and Integrator

The product/s has/have been assessed by application of the following standards or specifications:

# Low Voltage Directive (LVD) 2014/35/EU EN 61010-1:2010+A1:2019 Safety requirements for electrical equipment for measurement, control, and laboratory use. 600V CAT IV 1000V CAT III Pollution Degree 2 EN IEC 61010-2-032:2021+A11:2021 Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

### EMC Directive 2014/30/EU

### EN 61326-1:2013

Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements

Susceptibility Criteria A Emission Class B

### RoHS Directive 2011/65/EU & 2015/863/EU

Directive on the Restriction of the use of certain Hazardous Substances (RoHS) in Electrical and Electronic Equipment.

I certify for and on behalf of GMC-I PROSyS Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature	Warooli	Name	Robin Brooks
Position	R&D Director	Date	1st April 2025