

EUROPEAN UNION DECLARATION OF CONFORMITY

		We, the unde	rsigned,	
	Manu	facturer	Authorised Representative	
Name	Dranetz Technologies Inc.,		Eurolink Europe Compliance Ltd	
Address	191 Talmadge Road,		25 Herbert Place	
	Edison,		Dublin	
	NJ 08817		D02 AY86	
Country	USA		Republic of Ireland	
	certify and declare	under our sole respons	sibility that the following apparatus:	
Equipment Type:		Power Monitorin	Power Monitoring Equipment	
Apparatus Identification:		Model Number(s	Model Number(s)	
		HDPQ Xplor	HDPQ Xplorer, Xplorer 400, Guide, Visa	
		HDPQ Plus Xplor	HDPQ Plus Xplorer, Xplorer 400, Guide, Visa	
Directives Applied:			2014/30/EU Electromagnetic Compatibility	
		2014/35/EU Low	2014/35/EU Low Voltage Safety	
Harmonised Standards:		EN 61326: Class A	EN 61326: Class A, Electromagnetic Compatibility -	
			Referenced Electrical equipment for Measurement, Control and	
			Laboratory use	
			Standards CISPR 11: Class A Limits Of Radio Disturbance and Immunity Measuring Apparatus	
		EN 61000-4-2: Ele	EN 61000-4-2: Electrostatic Discharge Immunity	
		EN 61000-4-3: Ele	EN 61000-4-3: Electromagnetic Field Immunity	
		EN 61000-4-4: Ele	EN 61000-4-4: Electrical Fast Transient Burst Immunity	
		EN 61000-4-5: Su	EN 61000-4-5: Surge Immunity	
		EN 61000-4-6: Im	EN 61000-4-6: Immunity to Conducted Disturbances	
		EN 61010-1: Safe	EN 61010-1: Safety Requirements of Electrical Equipment for	
		Measurement, Co	Measurement, Control, and Laboratory Use	

- -Commich, cale

Signed: _______ Dominick Scala Engineering Product Support Manager Dranetz Technologies Inc.

Date of Issue: December 14, 2022

Issued: Dublin, Ireland

The Technical Documentation is kept at the Eurolink Europe Compliance Ltd offices, address shown above.



UNITED KINGDOM DECLARATION OF CONFORMITY

		We, the undersigr	ed,	
	Manufa	cturer	Authorised Representative	
Name	Dranetz Technologies Inc.	·,	Eurolink (Europe) Ltd	
Address	191 Talmadge Road, Edison, NJ 08817		Milestone House, Longcot Road, Shrivenham, Oxon	
Country	USA		UK	
	certify and declare un	der our sole responsibili	ty that the following apparatus:	
Equipment Type:		Power Monitoring Eq	Power Monitoring Equipment	
Apparatus Identification:		Model Number(s) HDPQ Xplorer, Xplorer 400, Guide, Visa HDPQ Plus Xplorer, Xplorer 400, Guide, Visa		
Directives Applied:		SI 2016 No. 1091 Electromagnetic Compatibility SI 2016 No. 1101 Electrical Safety		
Harmonised Standards:		EN 61326: Class A, Electromagnetic Compatibility -		
		Referenced Electrical equipment for Measurement, Control and Laboratory use		
		Standards CISPR 11: Class A Limits of Radio Disturbance and Immunity Measuring Apparatus		
		EN 61000-4-2: Electrostatic Discharge Immunity		
		EN 61000-4-3: Electromagnetic Field Immunity		
		EN 61000-4-4: Electrical Fast Transient Burst Immunity		
		EN 61000-4-5: Surge Immunity		
		EN 61000-4-6: Immunity to Conducted Disturbances		
		EN 61010-1: Safety Requirements of Electrical Equipment for Measurement, Control, and Laboratory Use		
and therefo	re complies with the essentia	al requirements and prov Regulations	isions of the UK EMC Regulations and Low Voltage	
	Signed: Jominich	Kaler	Date of Issue: December 14, 2022	
	Engineering Product Suppo Dranetz Technologies Inc.	ort Manager	Issued: Shrivenham, UK	
	The Technical Documentation	is kept at the Eurolink (Eu	rope) Ltd offices, address shown above.	